

**QUESTIONS AND ANSWERS N. 2****Reference is made to RFQ for Purchase of Dental Equipment In Support Of Women Entrepreneurship (UNOPS-EP-2016-G-001).****Question 1:** Autoclave for steam sterilization (LOT 2, item 1):

- Can the autoclave with total power of 1400 W be offered?
- Can the autoclave which has operating temperature from 121 to 134 be offered, considering that the temperature interval is wider than requested?
- Can the autoclave with sterilization pressure from 1,1 to 2,2, bars be offered, considering that interval is wider than requested?

**Answer 1:**

- Anything less than the required minimum, which is 1600 W, is not acceptable.
- If the offered wider range covers the minimum required range as stated in technical specifications  $128\text{ }^{\circ}\text{C} \pm 2\text{ }^{\circ}\text{C}$ , it is acceptable.
- Same as above.

**Question 2:** Ultrasonic Bath (LOT 2, item 3)

- Can a device with one ceramic heater be offered, considering that the number of heaters does not influence the functionality of the device?
- Can a device with selection of two washing frequencies and one control unit be offered, considering that the number of control units does not affect the quality of washing?
- Is it possible to offer a device that is not equipped with internal light, considering that is irrelevant for the device functioning?
- Is it possible to offer a machine without a cooling fan?

**Answer 2:**

- It is not acceptable, since two heaters are the required minimum. Any offer that doesn't comply with the required minimum technical characteristics is not acceptable.
- Any offer that doesn't comply with the required minimum technical characteristics is not acceptable.
- Same as above.
- Same as above.

**Question 3:** Intraoral Camera (LOT 5, item 1)

- Is it possible to offer a camera that is equipped with CMOS sensor 1/2.5 inch, which is a more advanced generation than CCD sensor?

**Answer 3:** The camera with CCD sensor is the minimum technical requirement for the oral camera. Bidders may offer items with more advanced specs and such offers are accepted as appropriate.

**Question 4:** Do the documents issued by authorities of the Republic of Serbia have to be notarized translation (by a court interpreter), or is it enough to have a free translation into English?

**Answer:** No translation of the documents is needed for the submission of quotation, thus documents in Serbian language are enough.

**Question 5:** Is it acceptable to separate the item 2 from LOT2 into a separate LOT, in order to enable the participation of bidders who in their sales programme have steam autoclaves as well as ultrasonic cleaners, but not dry sterilizers?

**Answer 5:** No partial quotations within the LOTs are allowed. In order for their offer to be considered, Bidders must offer 100 % of the items specified for for one or more lots identified in Section III and to 100% of the quantities specified for each item of a lot.

In Belgrade, 2 February 2016