

QUESTIONS AND ANSWERS

Reference is made to RFQ for Purchase of Dental Equipment In Support Of Women Entrepreneurship (UNOPS-EP-2016-G-001).

Question 1: Does Multi-Functional Foot Pedal (LOT 1, Item 1) shall have the possibility of programming the desired options, or just a function of activation of already programmed (preset) functions?

Answer 1: Multi-Functional Foot Pedal must have the possibility of programming the desired option and not only of choosing the preset option.

Question 2: Does Multi-Functional Foot Pedal (LOT 1, Item 1) must have the possibility to adjust intensity of reflector settings including individual activation?

Answer 2: Yes, Multi-Functional Foot Pedal must have the possibility of adjusting the intensity of reflector.

Question 3: Does Reflector (LOT 1, Item 1) need to have strictly color temperature from 5.500 – 6.300 K or one can offer color of 5.000 K?

Answer 3: 5,500 K is the required minimum, so 5,000 K is not acceptable.

Question 4: Does Reflector (LOT 1, Item 1) need to have strictly intensity adjustment from 10 – 35kLux or we can offer stronger intensity from 10 – 50kLux?

Answer 4: 10 – 35kLux is the required minimum, so stronger intensity range can also be offered and is acceptable.

Question 5: Regarding the installation of dental unit and preparation the connections in floor for factory needs, what are the obligations in this purchasing??

Answer 5: As communicated on the Notice published on 28 January 2015, installation at beneficiaries 'premises must be included. That means assembly and installation of the equipment to the existing electric system and fittings.

In Belgrade, 2 February 2016